

棒温度計°C	サーモ°C	誤差°C
10.0	10.9	0.9
15.0	15.4	0.4
20.0	19.5	-0.5
25.0	25.4	0.4
27.0	27.5	0.5
28.0	28.5	0.5
30.0	30.7	0.7
33.0	33.7	0.7
35.0	35.8	0.8
37.0	37.8	0.8
40.0	40.3	0.3
42.0	42.6	0.6
45.0	45.7	0.7
47.0	47.5	0.5
50.0	50.7	0.7
52.0	52.7	0.7
55.0	55.9	0.9
58.0	59.1	1.1
60.0	62.0	2.0
62.0	63.5	1.5
65.0	66.8	1.8
67.0	68.2	1.2
70.0	72.2	2.2
72.0	75.0	3.0
75.0	77.2	2.2
77.0	79.2	2.2
80.0	82.0	2.0
83.0	85.0	2.0
85.0	87.3	2.3

温度検証(棒温度計・サーモセンサ)

